



10B Airline Drive
Albany, New York 12235
800-554-4501
www.agriculture.ny.gov

Certificate of Approval
for Weighing and Measuring Devices

New York State Certificate Number: 9841
Effective Date: June 7, 2016

NTEP Certificate of Conformance Number: 16-031

For:

Non-Computing Scale
Digital Electronic
Model: PWT-xxK-M Series
 m_{max} : 3000
 e_{min} : (see table below)
Capacity: 3 kg to 30 kg (see table below)
Platform: 220 cm x 180 cm (8.5 in x 7 in)
Accuracy Class: III

Submitted By:

Nanjing Toms Weighing Instrument Co., Ltd.
No. 77 Baoshan Rd.
Qilin Town
Nanjing, Jiangsu 15206 China
Tel: +862587157179
Fax: +862584121983
Contact: Ming Tao
Email: sales-scale@tom168.com
Web site: www.toms168.com

This certifies that the items specified in the above National Type Evaluation Program (NTEP) Certificate of Conformance are hereby approved for sale or use in the State of New York.

The NTEP Certificate of Conformance, as issued by the National Conference on Weights and Measures, is based on an evaluation made in the New York State Laboratory. Evaluation results and device characteristics necessary for inspection and use in commerce are stated in the NTEP Certificate of Conformance. Copies of the NTEP Certificate of Conformance are available on request and are available for inspection at the Bureau's Metrology Office at 6 Harriman Campus Road, Albany, NY 12206.

Michael Sikula, Director
NYS Bureau of Weights and Measures



10B Airline Drive
Albany, New York 12235
800-554-4501
www.agriculture.ny.gov

Certificate of Approval
for Weighing and Measuring Devices

New York State Certificate Number: 9850
Effective Date: June 16, 2016

NTEP Certificate of Conformance Number: 16-046

For:

Non-Computing Scale
Digital Electronic
Model: EXxxxx Series
n_{max}: see table below
c_{mm}: see table below
Capacity: see table below
Platform: see table below
Accuracy Class: II

Submitted By:

Nanjing Toms Weighing Instrument Co., Ltd.
No. 77 Baoshan Rd.
Qilin Town
Nanjing, Jiangsu 15206 China
Tel: +862587157179
Fax: +862584121983
Contact: Ming Tao
Email: sales-scale@tom168.com
Web site: www.toms168.com

This certifies that the items specified in the above National Type Evaluation Program (NTEP) Certificate of Conformance are hereby approved for sale or use in the State of New York.

The NTEP Certificate of Conformance, as issued by the National Conference on Weights and Measures, is based on an evaluation made in the New York State Laboratory. Evaluation results and device characteristics necessary for inspection and use in commerce are stated in the NTEP Certificate of Conformance. Copies of the NTEP Certificate of Conformance are available on request and are available for inspection at the Bureau's Metrology Office at 6 Harriman Campus Road, Albany, NY 12206.

A handwritten signature in black ink, appearing to read "Michael Sikula".

Michael Sikula, Director
NYS Bureau of Weights and Measures